



Certificate of Calibration

Level Switch Measurement

Certificate No. : 0132-23-HI-CAL-101

Customer Information

Company Name PT. Bhumi Jati Power
Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
Contract No. C-BJP-23-0008-OTH

Instrument Information

KKS No. 50MAG15CL305 Type of Testing Function Test
Instrument Description CONDENSER WATER BOX A LEVEL HIGH Cal. Date 31-05-2023
Manufacture NAKAKITA
Model NS-287-71-A1F3-10L0G-F
Serial Number 6005
Instrument Location Turbin Lt. 1

Environment Condition

Temperature 30.1 °C
Humidity 70 % RH

Calibration Result

Data Reading (Switch)			
As Found		As Left	
Open to Close	Close to Open	Open to Close	Close to Open
OK	OK		
OK	OK		
OK	OK		

Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
ROLL METER	-	L04220787	26-04-2022
STEEL RULER 018/L3/HI/2020	-	S. 23 004 598	24-03-2023

Calibration Method

0008/SOP_BJS/2023
Procedure of Calibration Level Switch

This certificate comprise of 1 page
Date of issue, 01-06-2023

Customer Signature,

Ayrton Humardani Putradhi

Approved Sign **HERVITAMA**

INDONESIA

Arief Gunawan

Calibrated by,

R. Heru Setyo Utomo