



Report of Inspection

Vibration Sensor

Report No. : 0260-23-HI-CAL-10

Customer Information

Company Name PT. Bhumi Jati Power
Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
Contract No. C-BJP-23-0008-OTH

Instrument Information

KKS No. 50MAY10CY011 Inspection Date 19-05-2023
Instrument Description BEARING NO.6 VIBRATION X
Manufacture SHINKAWA
Model FL-202F08A-M2-00-05-10
Serial Number T9SH5269
Instrument Location Turbin Lt. 3

Environment Condition

Temperature 31 °C
Humidity 61 % RH

Inspection Result

Visual Inspection

Cabel Tightness	✓	OK		Not OK
Cable Condition	✓	OK		Not OK
Mounting	✓	OK		Not OK
Phisical Condition	✓	OK		Not OK

Healthiness

Good

Measurement

V - Supply	- 25.43
V - Out	- 09.04
Resistance [Ω]	8.7

Remark

Test Equipment

Description	Serial Number	Certificate Number	Calibration Date
TRUE RMS MULTIMETER	160403244	S. 21 011 088	05-10-2021

Inspection Method

0006/SOP_BJS/2023
Procedure of Inspection Vibration

This certificate comprise of 1 page
Date of issue, 20-05-2023

Customer Signature,

Ayrtan Humardani Putradhi

HERVITAMA
Approved Signatory,

Aris Nugroho

Calibrated by,

Abdul Latif