

## CERTIFICATE OF CALIBRATION LEVEL SWITCH

Certificate No. : 0491-24-HI-CAL-101

### Customer Information

Company Name PT. Bhumi Jati Power  
Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah  
Contract No. C-BJP-24-0013-0TH

### Instrument Information

KKS No. 60HFA60CL301 Type of Testing Function Test  
Instrument Description COAL SILO F LEVEL SWITCH 1 Cal. Date 19-05-2024  
Manufacture ENDRESS HAUSER  
Model FT156  
Serial Number P10003010FF  
Instrument Location BOILER 7TH FLOOR

### Environment Condition

Temperature 27.4 °C  
Humidity 51 % RH

### Calibration Result

Data Reading (Switch)			
As Found		As Left	
Switch	Result	Switch	Result
Set (NC)	OK		
Reset (NO)	OK		

### Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
Documenting Process Calibrator Fluke 754	56070038	S. 22 002 081	17-02-2022

### Calibration Method

0008/SOP\_BJS/2023  
Procedure of Calibration Level Switch

This certificate comprise of 1 page  
Date of issue, 20-05-2024

Customer Signature,




Ayrton Humardani Putradhi

Approved Signatory,



MINDONESIA Irian

Calibrated by,



Faisnol Anwar

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas  
This certificate applies only for the item specified above  
Dilarang mengutip, memperbanyak dan / atau mempublikasikan isi sertifikat ini tanpa izin dari PT. Hervitama Indonesia  
It is prohibited to copy, reproduce and / or publish the content of this Certificate without PT Hervitama Indonesia's approval  
\*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang  
This certificate is valid if it has approved and signed by authorized person

