

CERTIFICATE OF CALIBRATION TEMPERATURE SENSOR

Certificate No. : 0700-24-HI-CAL-10D

Customer Information

Company Name PT. Bhumi Jati Power
 Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
 Contract No. C-BJP-24-0013-OTH

Instrument Information

KKS No. 60HFE24CT002 Range 0 ~ 400
 Instrument Description PULVERIZER D PRIMARY AIR Unit °C
 TEMPERATURE 2
 Manufacture YAMARI Out. Signal 4 - 20 mA
 Model TM Max. Deviation 4 °C
 Serial Number 5451727 Cal. Date 11-05-2024
 Instrument Location BOILER 1ST FLOOR Temp. Element Thermocouple Type E

Environment Condition

Temperature 29.8 °C
 Humidity 50 % RH

Calibration Result

As Found					Data Reading					As Left				
Input	Std Output	Output UUT	Deviation	Result	Input	Std Output	Output UUT	Deviation	Result	Input	Std Output	Output UUT	Deviation	Result
[°C]	[°C]	[°C]			[°C]	[°C]	[°C]			[°C]	[°C]	[°C]		
40	40.000	40.400	0.400	Pass										
100	100.000	101.200	1.200	Pass										
200	200.000	202.500	2.500	Pass										
300	300.000	303.400	3.400	Pass										
400	400.000	403.700	3.700	Pass										

Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
Documenting Process Calibrator Fluke 754	43550006	S. 22 022 080	21-02-2022
DRYBLOCK ADDITEL 875	660TH190400016	S. 23 008 476 /P	03-07-2023

Calibration Method

0002/SOP_BJS/2023
 Procedure of Calibration Temperature Element and Temperature Transmitter

This certificate comprise of 1 page

Date of issue, 12-05-2024

Calibrated by,

Customer Signature,



Ayrton Humardani Putradhi

Approved Signatory,



M. Taufiq Nurfitriyan



Khabib Mundzirul Umam

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas

This certificate applies only for the item specified above

Dilarang mengutip, memperbanyak dan / atau mempublikasikan isi sertifikat ini tanpa izin dari PT. Hervitama Indonesia

It is prohibited to copy, reproduce and / or publish the content of this Certificate without PT Hervitama Indonesia's approval

*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang

This certificate is valid if it has approved and signed by authorized person

