

CERTIFICATE OF CALIBRATION TEMPERATURE SENSOR

Certificate No. : 0722-24-HI-CAL-10D

Customer Information

Company Name PT. Bhumi Jati Power
 Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
 Contract No. C-BJP-24-0013-OTH

Instrument Information

KKS No. 60HNA01CT003
 Instrument Description ECONOMIZER A OUTLET GAS TEMPERATURE 3
 Manufacture YAMARI
 Model -
 Serial Number 5451675
 Instrument Location BOILER 7TH FLOOR

Range 0 - 500
 Unit °C
 Out. Signal 4 - 20 mA
 Max. Deviation 5 °C
 Cal. Date 14-05-2024
 Temp. Element Thermocouple Type E

Environment Condition

Temperature 30.2 °C
 Humidity 58 % RH

Calibration Result

As Found					As Left				
Input	Std Output	Output UUT	Deviation	Result	Input	Std Output	Output UUT	Deviation	Result
[°C]	[°C]	[°C]			[°C]	[°C]	[°C]		
40	40.000	40.000	0.000	Pass					
125	125.000	125.900	0.900	Pass					
250	250.000	251.500	1.500	Pass					
375	375.000	375.700	0.700	Pass					
500	500.000	501.100	1.100	Pass					

Standard Reference


Description	Serial Number	Certificate Number	Calibration Date
Documenting Process Calibrator Fluke 754	43550006	S. 22 022 030	21-02-2022
DRYBLOCK ADDITEL 875	660TH190400016	S. 23 008 476 /P	03-07-2023

Calibration Method

0002/SOP_BJS/2023
 Procedure of Calibration Temperature Element and Temperature Transmitter

This certificate comprise of 1 page
 Date of issue, 15-05-2024

Customer Signature,



Ayrtan Humardani Putradhi

Approved Signatory,



Khabib Mundzirul Umam

Calibrated by,



Khabib Mundzirul Umam

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas

This certificate applies only for the item specified above

Dilarang mengutip, memperbanyak dan / atau mempublikasikan isi sertifikat ini tanpa izin dari PT. Hervitama Indonesia

It is prohibited to copy, reproduce and / or publish the content of this Certificate without PT Hervitama Indonesia's approval

*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang

This certificate is valid if it has approved and signed by authorized person

