

## CERTIFICATE OF CALIBRATION PRESSURE TRANSMITTER

Certificate No. : 0473-24-HI-CAL-10C

### Customer Information

Company Name: PT. Bhumi Jati Power  
Address: PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah  
Contract No.: C-BJP-24-0013-OTH

### Instrument Information

KKS No.	60HTB20CP003	Range	0 ~ 25
Instrument Description	U6 SOOT BLOWER STEAM INLET PRESSURE	Unit	bar
Manufacture	ROSEMOUNT	Out. Signal	4 - 20 mA
Model	64HANAXAJ6M4Q4	Max. Error	0.5 %
Serial Number	3666972	Cal. Date	24-05-2024
Instrument Location	GGH		

### Environment Condition

Temperature: 32.3 °C  
Humidity: 65 % RH

### Calibration Result

Data Reading (Transmitter)											
As Found						As Left					
Input	Std Output	Output UUT	Deviation	Error	Result	Input	Std Output	Output UUT	Deviation	Error	Result
[bar]	[mA]	[mA]		[%]		[bar]	[mA]	[mA]		[%]	
0.000	4.000	4.000	0.000	0.000	Pass						
6.254	8.003	8.008	0.005	0.031	Pass						
12.500	12.000	12.007	0.007	0.044	Pass						
18.754	16.003	16.008	0.005	0.031	Pass						
25.000	20.000	20.003	0.003	0.019	Pass						
18.754	16.003	16.008	0.005	0.031	Pass						
12.500	12.000	12.002	0.002	0.013	Pass						
6.254	8.003	8.005	0.002	0.013	Pass						
0.000	4.000	4.000	0.000	0.000	Pass						

### Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
Documenting Process Calibrator Fluke 754	3611011	S. 22 011 747	23-09-2023
Fluke Pressure Module 700P08	14070183	S. 23 004 594	27-03-2023

### Calibration Method

0001/SOP\_BJS/2023  
Procedure of Calibration Pressure & Differential Transmitter

This certificate comprise of 1 page  
Date of issue, 25-05-2024

Customer Signature,



Ayrlton Humardani Putradhi

HERVITAMA  
Approved Signatory,



M. Tjandjaja Nurfitriyan

Calibrated by,



Handi Suryawinata

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas

*This certificate applies only for the item specified above*

Dilarang mengutip, memperbanyak dan / atau mempublikasikan isi sertifikat ini tanpa izin dari PT. Hervitama Indonesia

*It is prohibited to copy, reproduce and / or publish the content of this Certificate without PT Hervitama Indonesia's approval*

\*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang

*This certificate is valid if it has approved and signed by authorized person*

