

CERTIFICATE OF CALIBRATION TEMPERATURE SENSOR

Certificate No. : 0583-24-HI-CAL-10D

Customer Information

Company Name PT. Bhumi Jati Power
 Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
 Contract No. C-BJP-24-0013-OTH

Instrument Information

KKS No.	60LAB01CT005	Range	0 ~ 250
Instrument Description	TURBINE DRIVEN BOILER FEEDWATER PUMP A INLET WATER TEMPERATURE	Unit	°C
Manufacture	YAMARI	Out. Signal	4 - 20 mA
Model	-	Max. Deviation	2.5 °C
Serial Number	5346931	Cal. Date	17-05-2024
Instrument Location	TURBINE 1ST FLOOR	Temp. Element	Thermocouple Type E

Environment Condition

Temperature 27.4 °C
 Humidity 49 % RH

Calibration Result

As Found					Data Reading					As Left				
Input	Std Output	Output UUT	Deviation	Result	Input	Std Output	Output UUT	Deviation	Result	Input	Std Output	Output UUT	Deviation	Result
[°C]	[°C]	[°C]			[°C]	[°C]	[°C]			[°C]	[°C]	[°C]		
40.0	40.010	40.100	0.090	Pass										
62.5	62.500	63.000	0.500	Pass										
125.0	125.020	126.300	1.280	Pass										
187.5	187.480	189.400	1.920	Pass										
250.0	250.000	252.200	2.200	Pass										

Standard Reference


Description	Serial Number	Certificate Number	Calibration Date
Documenting Process Calibrator Fluke 754	43550006	S. 22 022 080	21-02-2022
DRYBLOCK ADDITEL 875	660TH190400016	S. 23 008 476 /P	03-07-2023

Calibration Method

0002/SOP_BJS/2023
 Procedure of Calibration Temperature Element and Temperature Transmitter

This certificate comprise of 1 page
 Date of issue, 18-05-2024

Customer Signature,




Ayrton Humardani Putradhi

Approved Signatory,
 HERVITAMA



M. INDONESIA

Calibrated by,



Khabil Mundzirul Umam

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas

This certificate applies only for the item specified above

Dilarang mengutip, memperbanyak dan / atau mempublikasikan isi sertifikat ini tanpa izin dari PT. Hervitama Indonesia
 It is prohibited to copy, reproduce and / or publish the content of this Certificate without PT Hervitama Indonesia's approval

*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang
 This certificate is valid if it has approved and signed by authorized person

