

CERTIFICATE OF CALIBRATION
LEVEL TRANSMITTER

Certificate No. : 0418-26-HI-CAL-10

Customer Information

Company Name PT. Bhumi Jati Power
Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
Contract No. C-BJP-26-0004-OTH

Instrument Information

KKS No. 60LAD32CL010 Range -250 ~ 660
Instrument Description HP HEATER 8B LEVEL 2 Unit mm
Manufacture YOKOGAWA Out. Signal 4 - 20 mA
Model / Serial Number EJX118A / 91V7003559377 Max. Error 0.5%
Instrument Location TURBINE 3RD FLOOR Cal. Date 06-06-2026

Environment Condition

Temperature 26 °C Humidity 64 % RH

Calibration Result

Data Reading (Transmitter)													
As Found							As Left						
Input	Std Output	Output UUT	Output UUT	Deviation	Error	Result	Input	Std Output	Output UUT	Output UUT	Deviation	Error	Result
[mm]	[mm]	[mm]	[mA]		[%]		[mm]	[mm]	[mm]	[mA]		[%]	
250	-204	-221	-	-17.000	-1.868	Fail	-250	-239	-239	4.181	0.000	0.000	Pass
-22.5	-13	-23	-	-10.000	-1.099	Fail	-22.5	-16	-16	8.112	0.000	0.000	Pass
205	198	187	-	-11.000	-1.209	Fail	205	185	184	16.637	-1.000	-0.110	Pass
432.5	434	424	-	-10.000	-1.099	Fail	432.5	418	419	15.773	1.000	0.110	Pass
660	658	645	-	-13.000	-1.429	Fail	660	659	661	20.011	2.000	0.220	Pass

Visual Inspection (Checklist if OK)

Leak Check ☒ Tightness Check ☒ Corrosion Check ☒
Accessories Check ☒ Power Supply ☒

Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
DIGITAL LASER DISTANCE METER	18H114632	02169/GQI-Sert/03/26	09-03-2026
STEEL TAPE MEASURE	001/L1/HI/2026	01231/GQI-Sert/02/26	09-02-2026

Calibration Method

0005/SOP_BJS/2025
Standard Operating Procedure of Calibration Level Transmitter Displacer

This certificate comprises 2 pages
Date of issue, 10-06-2026

Customer Signature,

Ayrton Humardani Putradhi

Approved by,

Ivan Hasan Ashari

Calibrated by,

Faishol Anwar

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas

This certificate applies only for the item specified above

Tidak dibenarkan mempublikasikan/memperbanyak sebagian isi sertifikat ini tanpa ijin PT. Hervitama Indonesia

This certificate shall not be produced/reproduced except in full without the written approval of PT. Hervitama Indonesia

*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang

This certificate is valid if it has approved and signed by authorized person