

CERTIFICATE OF CALIBRATION TEMPERATURE SENSOR

Certificate No. : 0606-24-HI-CAL-10D

Customer Information

Company Name PT. Bhumi Jati Power
 Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
 Contract No. C-BJP-24-0013-OTH

Instrument Information

KKS No. 60LAV21CT005 Range 0 ~ 100
 Instrument Description BFPT A MAIN OIL COOLER OUTLET OIL TEMPERATURE 1 Unit °C
 Manufacture YAMARI Out. Signal 4 - 20 mA
 Model - Max. Deviation 1.7 °C (JIS)
 Serial Number 5346944 Cal. Date 17-05-2024
 Instrument Location TURBINE UNDER GROUND Temp. Element Thermocouple Type E

Environment Condition

Temperature 27.6 °C
 Humidity 50 % RH

Calibration Result

Data Reading									
As Found					As Left				
Input	Std Output	Output UUT	Deviation	Result	Input	Std Output	Output UUT	Deviation	Result
[°C]	[°C]	[°C]			[°C]	[°C]	[°C]		
40	40.010	40.100	0.090	Pass					
45	45.010	45.200	0.190	Pass					
50	50.030	50.200	0.170	Pass					
75	75.020	75.600	0.580	Pass					
100	100.020	100.800	0.780	Pass					

Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
Documenting Process Calibrator Fluke 754	43550006	S. 22 022 080	21-02-2022
DRYBLOCK ADDITEL 875	660TH190400016	S. 23 008 476 /P	03-07-2023

Calibration Method

0002/SOP_BJS/2023
 Procedure of Calibration Temperature Element and Temperature Transmitter

Customer Signature,



Ayrton Humardani Putradhi

Approved Signatory,



M. Indro Nur Syahyan

Calibrated by,



Khabib Mundzirul Umam

This certificate comprise of 1 page
 Date of issue, 18-05-2024

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas
 This certificate applies only for the item specified above
 Dilarang mengutip, memperbanyak dan / atau mempublikasikan isi sertifikat ini tanpa izin dari PT. Hervitama Indonesia
 It is prohibited to copy, reproduce and / or publish the content of this Certificate without PT Hervitama Indonesia's approval
 *Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang
 This certificate is valid if it has approved and signed by authorized person

