

CERTIFICATE OF CALIBRATION TEMPERATURE SENSOR

Certificate No. : 0618-24-HI-CAL-10D

Customer Information

Company Name PT. Bhumi Jati Power
Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
Contract No. C-BJP-24-0013-OTH

Instrument Information

| | | | |
|------------------------|--|----------------|---------------------|
| KKS No. | 60LBR10CT005 | Range | 0 ~ 400 |
| Instrument Description | BFPT INLET STEAM HEADER TEMPERATURE | Unit | °C |
| Manufacture | YAMARI | Out. Signal | 4 - 20 mA |
| Model | - | Max. Deviation | 4 °C |
| Serial Number | - | Cal. Date | 19-05-2024 |
| Instrument Location | TURBINE 2ND FLOOR | Temp. Element | Thermocouple Type E |

Environment Condition

Temperature 27.7 °C
Humidity 57 % RH

Calibration Result

| Data Reading | | | | | | | | | |
|--------------|------------|------------|-----------|--------|---------|------------|------------|-----------|--------|
| As Found | | | | | As Left | | | | |
| Input | Std Output | Output UUT | Deviation | Result | Input | Std Output | Output UUT | Deviation | Result |
| [°C] | [°C] | [°C] | | | [°C] | [°C] | [°C] | | |
| 40 | 40.000 | 40.100 | 0.100 | Pass | | | | | |
| 100 | 100.000 | 101.000 | 1.000 | Pass | | | | | |
| 200 | 199.990 | 201.100 | 1.110 | Pass | | | | | |
| 300 | 300.020 | 301.100 | 1.080 | Pass | | | | | |
| 400 | 400.000 | 401.200 | 1.200 | Pass | | | | | |

Standard Reference

| Description | Serial Number | Certificate Number | Calibration Date |
|----------------------------------|----------------|--------------------|------------------|
| DRYBLOCK ADDITEL 875 | 660TH190400016 | S. 23 008 476 /P | 03-07-2023 |
| Temperature Calibrator Fluke 724 | 5534569 | S. 23 004 603 | 17-04-2023 |

Calibration Method

0002/SOP_BJS/2023
Procedure of Calibration Temperature Element and Temperature Transmitter

This certificate comprise of 1 page
Date of issue, 20-05-2024

Customer Signature,




Ayrton Humardani Putradhi

Approved Signatory,



M. Taufiq Nur Fitriyan

Calibrated by,



Khabib Mundzirul Umam

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas

This certificate applies only for the item specified above

Dilarang mengutip, memperbanyak dan / atau mempublikasikan isi sertifikat ini tanpa izin dari PT. Hervitama Indonesia
It is prohibited to copy, reproduce and / or publish the content of this Certificate without PT Hervitama Indonesia's approval

*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang

This certificate is valid if it has approved and signed by authorized person

