

**CERTIFICATE OF CALIBRATION
LEVEL TRANSMITTER**

Certificate No. : 1128-26-HI-CAL-10

Customer Information

Company Name PT. Bhumi Jati Power
Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
Contract No. C-BJP-26-0004-OTH

Instrument Information

KKS No.	60LCC30CL010	Range	-825 ~ 25
Instrument Description	LP HEATER 3 LEVEL 2	Unit	mm
Manufacture	TOKYO KEISO	Out. Signal	4 - 20 mA
Model / Serial Number	FST 4412 / L19-10464-3-2	Max. Error	0.5%
Instrument Location	TURBINE 2ND FLOOR	Cal. Date	03-06-2026

Environment Condition

Temperature 30 °C Humidity 84 % RH

Calibration Result

Data Reading (Transmitter)													
As Found							As Left						
Input	Std Output	Output UUT	Output UUT	Deviation	Error	Result	Input	Std Output	Output UUT	Output UUT	Deviation	Error	Result
[mm]	[mm]	[mm]	[mA]		[%]		[mm]	[mm]	[mm]	[mA]		[%]	
-825	-803	-826	3.981	-23.000	-2.706	Fail	-825	-799	-801	4.452	-2.000	-0.235	Pass
-612.5	-528	-548	9.214	-20.000	-2.353	Fail	-612.5	-598	-601	8.216	-3.000	-0.353	Pass
-400	-344	-364	12.678	-20.000	-2.353	Fail	-400	-393	-394	12.113	-1.000	-0.118	Pass
-187.5	-184	-204	15.689	-20.000	-2.353	Fail	-187.5	-183	-183	16.085	0.000	0.000	Pass
25	24	5	19.624	-19.000	-2.235	Fail	25	25	26	20.019	1.000	0.118	Pass

Visual Inspection (Checklist if OK)

Leak Check	☒	Tightness Check	☒	Corrosion Check	☒
Accessories Check	☒	Power Supply	☒		

Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
DIGITAL LASER DISTANCE METER	18H114632	02169/GQI-Sert/03/26	09-03-2026
STEEL TAPE MEASURE	001/L1/HI/2026	01231/GQI-Sert/02/26	09-02-2026

Calibration Method

0005/SOP_BJS/2025
Standard Operating Procedure of Calibration Level Transmitter Displacer

Customer Signature,

Ayrtan Humardani Putradhi

HERVITAMA
Approved Signatory,

INDONESIA
Ivan Hasan Ashari

This certificate comprises 2 pages
Date of issue, 08-06-2026

Calibrated by,

Faisho Anwar

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas

This certificate applies only for the item specified above

Tidak dibenarkan mempublikasikan/memperbanyak sebagian isi sertifikat ini tanpa ijin PT. Hervitama Indonesia

This certificate shall not be produced/reproduced except in full without the written approval of PT. Hervitama Indonesia

*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang

This certificate is valid if it has approved and signed by authorized person

