

CERTIFICATE OF CALIBRATION
LEVEL TRANSMITTER

Certificate No. : 1137-26-HI-CAL-10

Customer Information

Company Name PT. Bhumi Jati Power
Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah
Contract No. C-BJP-26-0004-OTH

Instrument Information

KKS No. 60MAG10CL015 Range -700 ~ 500
Instrument Description CONDENSER HOT WELL LEVEL 3 Unit mm
Manufacture TOKYO KEISO Out. Signal 4 - 20 mA
Model / Serial Number FST 4000 / L19-10464-6-3 Max. Error 0.5%
Instrument Location TURBINE UNDERGROUND Cal. Date 08-06-2026

Environment Condition

Temperature 27 °C Humidity 74 % RH

Calibration Result

Data Reading (Transmitter)													
As Found							As Left						
Input	Std Output	Output UUT	Output UUT	Deviation	Error	Result	Input	Std Output	Output UUT	Output UUT	Deviation	Error	Result
[mm]	[mm]	[mm]	[mA]		[%]		[mm]	[mm]	[mm]	[mA]		[%]	
-700	-641	-633	-	8.000	0.667	Fail	-700	-676	-677	4.311	-1.000	-0.083	Pass
-400	-426	-416	-	10.000	0.833	Fail	-400	-422	-423	7.687	-1.000	-0.083	Pass
-100	-110	-107	-	3.000	0.250	Pass	-100	-79	-80	12.267	-1.000	-0.083	Pass
200	206	208	-	2.000	0.167	Pass	200	186	183	15.776	-3.000	-0.250	Pass
500	498	497	-	-1.000	-0.083	Pass	500	500	497	19.967	-3.000	-0.250	Pass

Visual Inspection (Checklist if OK)

Leak Check ☒ Tightness Check ☒ Corrosion Check ☒
Accessories Check ☒ Power Supply ☒

Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
DIGITAL LASER DISTANCE METER	18H114632	02169/GQI-Ser/03/26	09-03-2026
STEEL TAPE MEASURE	001/L1/HI/2026	01231/GQI-Ser/02/26	09-02-2026

Calibration Method

0005/SOP_BJS/2025
Standard Operating Procedure of Calibration Level Transmitter Displacer

This certificate comprises 2 pages
Date of issue, 12-06-2026

Customer Signature,

Ayrtan Humardani Putradhi

Approved Signature,

Ivan Hasan Ashari

Calibrated by,

Faishol Anwar

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas

This certificate applies only for the item specified above

Tidak dibenarkan mempublikasikan/memperbanyak sebagian isi sertifikat ini tanpa ijin PT. Hervitama Indonesia

This certificate shall not be produced/reproduced except in full without the written approval of PT. Hervitama Indonesia

*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang

This certificate is valid if it has approved and signed by authorized person

