

## CERTIFICATE OF CALIBRATION PRESSURE TRANSMITTER

Certificate No. : 0403-24-HI-CAL-10C

### Customer Information

Company Name PT. Bhumi Jati Power  
 Address PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah  
 Contract No. C-BJP-24-0013-OTH

### Instrument Information

KKS No. 60MAV83CP002  
 Instrument Description JACKING OIL PUMP B INLET OIL PRESSURE  
 Manufacture YOKOGAWA  
 Model EJX430A  
 Serial Number 91V510234  
 Instrument Location TURBINE 2ND FLOOR

Range 0 ~ 10  
 Unit barg  
 Out. Signal 4 - 20 mA  
 Max. Error 0.5 %  
 Cal. Date 15-05-2024

### Environment Condition

Temperature 30.7 °C  
 Humidity 59 % RH

### Calibration Result

As Found						As Left					
Input	Std Output	Output UUT	Deviation	Error	Result	Input	Std Output	Output UUT	Deviation	Error	Result
[barg]	[mA]	[mA]		[%]		[barg]	[mA]	[mA]		[%]	
0.0	4.000	3.993	-0.007	-0.044	Pass						
2.5	8.000	7.997	-0.003	-0.019	Pass						
5.0	12.000	11.989	-0.011	-0.069	Pass						
7.5	16.000	15.984	-0.016	-0.100	Pass						
10.0	20.000	19.975	-0.025	-0.156	Pass						
10.0	20.000	19.975	-0.025	-0.156	Pass						
7.5	16.000	15.980	-0.020	-0.125	Pass						
5.0	12.000	11.981	-0.019	-0.119	Pass						
2.5	8.000	8.012	0.012	0.075	Pass						
0.0	4.000	4.025	0.025	0.156	Pass						

### Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
Documenting Process Calibrator Fluke 754	3611011	S. 22 011 747	23-09-2023
Fluke Pressure Module 750PD27	6087359	S. 23 014 459	09-10-2023

### Calibration Method

0001/SOP\_BJS/2023  
 Procedure of Calibration Pressure & Differential Transmitter

This certificate comprise of 1 page  
 Date of issue, 16-05-2024

Customer Signature,



Ayrton Humardani Putradhi

Approve / Signatory,



M. Laili Nur Hidayat

INDONESIA

Calibrated by,



Handi Suryawinata

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas  
 This certificate applies only for the item specified above  
 Dilarang mengutip, memperbanyak dan / atau mempublikasikan isi sertifikat ini tanpa izin dari PT. Hervitama Indonesia  
 It is prohibited to copy, reproduce and / or publish the content of this Certificate without PT Hervitama Indonesia's approval  
 \*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang  
 This certificate is valid if it has approved and signed by authorized person