

## CERTIFICATE OF CALIBRATION DIFFERENTIAL PRESSURE TRANSMITTER

Certificate No. : 0408-24-HI-CAL-10C

### Customer Information

Company Name: PT. Bhumi Jati Power  
Address: PLTU Tanjung Jati B Unit 5&6 Tubanan, Kembang, Jepara - Jawa Tengah  
Contract No.: C-BJP-24-0013-OTH

### Instrument Information

KKS No.	60MKW30CP001	Range	0 ~ 1150
Instrument Description	GENERATOR SEAL OIL DIFFERENTIAL PRESSURE	Unit	mbar
Manufacture	YOKOGAWA	Out. Signal	4 - 20 mA
Model	EJX413A	Max. Error	0.5 %
Serial Number	91VA01021	Cal. Date	28-05-2024
Instrument Location	TURBINE 1ST FLOOR		

### Environment Condition

Temperature: 30 °C  
Humidity: 77 % RH

### Calibration Result

Data Reading (Transmitter)											
As Found						As Left					
Input	Std Output	Output UUT	Deviation	Error	Result	Input	Std Output	Output UUT	Deviation	Error	Result
[mbar]	[mA]	[mA]		[%]		[mbar]	[mA]	[mA]		[%]	
150	4.000	4.006	0.006	0.037	Pass						
400	8.000	8.021	0.021	0.131	Pass						
650	12.000	11.994	-0.006	-0.037	Pass						
900	16.000	15.973	-0.027	-0.169	Pass						
1150	20.000	19.977	-0.023	-0.144	Pass						
1150	20.000	19.977	-0.023	-0.144	Pass						
900	16.000	15.975	-0.025	-0.156	Pass						
650	12.000	12.007	0.007	0.044	Pass						
400	8.000	7.995	-0.005	-0.031	Pass						
150	4.000	4.002	0.002	0.013	Pass						

### Standard Reference

Description	Serial Number	Certificate Number	Calibration Date
Documenting Process Calibrator Fluke 754	3611011	S. 22 011 747	23-09-2023
PRESSURE MODULE 700 P05	14270508	S. 22 011 744	20-11-2022

### Calibration Method

0001/SOP\_BJS/2023  
Procedure of Calibration Pressure & Differential Transmitter

This certificate comprise of 1 page  
Date of issue, 29-05-2024

Customer Signature,



Ayrton Humardani Putradhi

HERVITAMA  
Approved Signatory,



M. Taufiq Nur Fitriyan

Calibrated by,



Handi Suryawinata

Sertifikat ini berlaku untuk alat dengan nama, merek dan tipe / nomor seri yang dinyatakan diatas

*This certificate applies only for the item specified above*

Dilarang mengutip, memperbanyak dan / atau mempublikasikan isi sertifikat ini tanpa izin dari PT. Hervitama Indonesia

*It is prohibited to copy, reproduce and / or publish the content of this Certificate without PT Hervitama Indonesia's approval*

*\*Sertifikat ini sah bila dibubuhi cap PT. Hervitama Indonesia dan ditandatangani oleh pejabat yang berwenang*

*This certificate is valid if it has approved and signed by authorized person*

